

Certificate of Conformance**SMARTCHECK
MPI BLACK WB
MPIBWB1**

Certificate Number: C0455
Batch: 0423594
Date of Manufacture: 11/04/2023
Expiry: 11/04/2026

The above batch of SMARTCHECK MPI BLACK WB has been tested and conforms to the following specifications:-

AMS 3042	ASTM E-3024
AS 1171	ISO 9934-2

Results per ISO 9934-2

	Result	Limits
7.1 Performance (ref Block 1)	Pass	Equal to Std
7.2 Colour	Pass	Equal to Std
7.3 Particle size (µm)		Report
	d1 (10%)	0.2
	da (50%)	1.1
	du (10%)	4.1
7.13 Storage Stability	11/04/2026	3 years from DOM

Signed: 

Date: 11/04/2023